



# Data Report

File Created Date: 30-Nov-21 14:38:05

Note: All times shown are based on UTC +00:00 and 24-Hour clock [DD-MMM-YY HH:MM:SS]

## Device Information

Serial Number: S2H21050014-17	Log Interval: 00H 01M 00S	Temperature Type: °C
Model Number: N/A	Start Mode: Manual	Light Detect: Enable
Version: V1.0	Start Delay: 00D 00H 00M 00S	

## Trip Information

Internal ID: 0000001  
 Description: TEST

## Logging Summary

Highest Temperature: 24.5 °C / 75.6 %	Record Mode: Stop when full	Start Time: 12-Nov-21 13:25:27
Lowest Temperature: 11.7 °C / 33.3 %	Stop Mode: Temporary	Stop Time: 30-Nov-21 14:37:27
Average Temperature: 18.7 °C / 51.4 %	Data Points: 25993	Elapsed Time: 18D 01H 12M 00S
MKT: 19.0 °C	First Alarm(T): 12-Nov-21 13:50:27	First Alarm(rH): 12-Nov-21 13:25:27


## Marked Events

N/A

Alarm Zone	Alarm Delay	Total Time	No. of Violations	Status
H3: Not Set				
H2: Not Set				
H1: over 24.0 °C	00D 00H 00M 00S (Sin)	00D 00H 16M 00S	3	Alarm
SZ: 24.0 to 15.0 °C	unlimited	16D 09H 17M 00S		
L1: below 15.0 °C	00D 00H 00M 00S (Sin)	01D 15H 40M 00S	12	Alarm
L2: Not Set				
HH: over 52.0 %	00D 00H 00M 00S (Sin)	09D 12H 27M 00S	104	Alarm
SZ: 52.0 to 15.0 %	unlimited	08D 12H 46M 00S		
HL: below 15.0 %	00D 00H 00M 00S (Sin)	00D 00H 00M 00S	0	OK

## Temperature (°C)

Temperature Humidity Mark Error Light Up Humidity (%RH)

































DATE TIME	°C	%RH	DATE TIME	°C	%RH	DATE TIME	°C	%RH	DATE TIME	°C	%RH	DATE TIME	°C	%RH	DATE TIME	°C	%RH
*18-Nov-21 09:25:27	20.8	50.7	*18-Nov-21 11:05:27	22.2	49.1	*18-Nov-21 12:45:27	22.8	44.1	*18-Nov-21 14:25:27	23.3	44.2	18-Nov-21 16:05:27	21.5	48.3	18-Nov-21 17:45:27	21.3	49.0
*18-Nov-21 09:26:27	20.8	50.8	*18-Nov-21 11:06:27	22.2	49.0	*18-Nov-21 12:46:27	22.8	44.2	*18-Nov-21 14:26:27	23.2	43.5	18-Nov-21 16:06:27	21.4	48.4	18-Nov-21 17:46:27	21.3	49.1
*18-Nov-21 09:27:27	20.8	50.8	*18-Nov-21 11:07:27	22.2	49.0	*18-Nov-21 12:47:27	22.8	44.0	*18-Nov-21 14:27:27	23.2	44.2	18-Nov-21 16:07:27	21.4	48.4	18-Nov-21 17:47:27	21.3	49.0
*18-Nov-21 09:28:27	20.8	50.7	*18-Nov-21 11:08:27	22.2	48.7	*18-Nov-21 12:48:27	22.8	44.0	*18-Nov-21 14:28:27	23.2	43.0	18-Nov-21 16:08:27	21.4	48.6	18-Nov-21 17:48:27	21.3	48.9
*18-Nov-21 09:29:27	20.9	50.6	*18-Nov-21 11:09:27	22.3	46.8	*18-Nov-21 12:49:27	22.8	44.0	*18-Nov-21 14:29:27	23.2	43.3	18-Nov-21 16:09:27	21.4	48.5	18-Nov-21 17:49:27	21.3	48.9
*18-Nov-21 09:30:27	20.9	50.7	*18-Nov-21 11:10:27	22.2	47.6	*18-Nov-21 12:50:27	22.8	43.9	*18-Nov-21 14:30:27	23.1	43.2	18-Nov-21 16:10:27	21.4	48.6	18-Nov-21 17:50:27	21.3	49.0
*18-Nov-21 09:31:27	20.9	50.7	*18-Nov-21 11:11:27	22.1	47.4	*18-Nov-21 12:51:27	22.8	44.9	*18-Nov-21 14:31:27	23.1	43.4	18-Nov-21 16:11:27	21.3	48.8	18-Nov-21 17:51:27	21.3	49.0
*18-Nov-21 09:32:27	20.9	50.6	*18-Nov-21 11:12:27	22.2	48.0	*18-Nov-21 12:52:27	22.8	45.8	*18-Nov-21 14:32:27	23.1	43.4	18-Nov-21 16:12:27	21.3	48.7	18-Nov-21 17:52:27	21.3	49.0
*18-Nov-21 09:33:27	20.9	50.7	*18-Nov-21 11:13:27	22.2	48.3	*18-Nov-21 12:53:27	22.9	46.1	*18-Nov-21 14:33:27	23.1	43.5	18-Nov-21 16:13:27	21.3	48.6	18-Nov-21 17:53:27	21.3	49.0
*18-Nov-21 09:34:27	20.9	50.7	*18-Nov-21 11:14:27	22.2	47.8	*18-Nov-21 12:54:27	22.9	45.2	*18-Nov-21 14:34:27	23.1	43.5	18-Nov-21 16:14:27	21.4	48.8	18-Nov-21 17:54:27	21.2	49.1
*18-Nov-21 09:35:27	20.8	50.7	*18-Nov-21 11:15:27	22.2	47.6	*18-Nov-21 12:55:27	23.0	45.2	*18-Nov-21 14:35:27	23.1	43.2	18-Nov-21 16:15:27	21.4	48.6	18-Nov-21 17:55:27	21.3	48.9
*18-Nov-21 09:36:27	20.9	50.6	*18-Nov-21 11:16:27	22.3	47.7	*18-Nov-21 12:56:27	23.0	44.9	*18-Nov-21 14:36:27	23.1	44.3	18-Nov-21 16:16:27	21.4	48.6	18-Nov-21 17:56:27	21.3	48.9
*18-Nov-21 09:37:27	20.9	50.7	*18-Nov-21 11:17:27	22.2	48.0	*18-Nov-21 12:57:27	23.0	45.1	*18-Nov-21 14:37:27	23.1	44.4	18-Nov-21 16:17:27	21.4	48.5	18-Nov-21 17:57:27	21.2	48.8
*18-Nov-21 09:38:27	20.9	50.6	*18-Nov-21 11:18:27	22.3	49.4	*18-Nov-21 12:58:27	23.0	44.3	*18-Nov-21 14:38:27	23.1	44.3	18-Nov-21 16:18:27	21.4	48.6	18-Nov-21 17:58:27	21.3	48.9
*18-Nov-21 09:39:27	20.9	50.7	*18-Nov-21 11:19:27	22.2	47.3	*18-Nov-21 12:59:27	23.0	44.9	*18-Nov-21 14:39:27	23.1	44.2	18-Nov-21 16:19:27	21.3	48.6	18-Nov-21 17:59:27	21.2	48.9
*18-Nov-21 09:40:27	20.9	50.6	*18-Nov-21 11:20:27	22.3	48.0	*18-Nov-21 13:00:27	23.0	45.0	*18-Nov-21 14:40:27	23.1	44.1	18-Nov-21 16:20:27	21.3	48.5	18-Nov-21 18:00:27	21.3	48.9
*18-Nov-21 09:41:27	20.9	50.6	*18-Nov-21 11:21:27	22.3	47.9	*18-Nov-21 13:01:27	23.0	44.9	*18-Nov-21 14:41:27	23.2	44.3	18-Nov-21 16:21:27	21.3	48.6	18-Nov-21 18:01:27	21.3	48.9
*18-Nov-21 09:42:27	20.9	50.5	*18-Nov-21 11:22:27	22.3	46.3	*18-Nov-21 13:02:27	23.0	44.7	*18-Nov-21 14:42:27	23.1	44.0	18-Nov-21 16:22:27	21.3	48.7	18-Nov-21 18:02:27	21.2	48.9
*18-Nov-21 09:43:27	21.0	50.5	*18-Nov-21 11:23:27	22.2	46.2	*18-Nov-21 13:03:27	23.1	43.4	*18-Nov-21 14:43:27	23.1	44.2	18-Nov-21 16:23:27	21.2	48.6	18-Nov-21 18:03:27	21.2	48.9
*18-Nov-21 09:44:27	21.0	50.5	*18-Nov-21 11:24:27	22.3	47.5	*18-Nov-21 13:04:27	23.1	44.7	*18-Nov-21 14:44:27	23.1	44.7	18-Nov-21 16:24:27	21.2	48.6	18-Nov-21 18:04:27	21.2	48.9
*18-Nov-21 09:45:27	21.0	50.6	*18-Nov-21 11:25:27	22.2	46.4	*18-Nov-21 13:05:27	23.1	44.3	*18-Nov-21 14:45:27	23.1	44.9	18-Nov-21 16:25:27	21.2	48.5	18-Nov-21 18:05:27	21.2	48.9
*18-Nov-21 09:46:27	21.0	50.5	*18-Nov-21 11:26:27	22.2	47.1	*18-Nov-21 13:06:27	23.1	43.6	*18-Nov-21 14:46:27	23.2	44.7	18-Nov-21 16:26:27	21.2	48.4	18-Nov-21 18:06:27	21.2	48.8
*18-Nov-21 09:47:27	20.9	50.5	*18-Nov-21 11:27:27	22.2	47.6	*18-Nov-21 13:07:27	23.1	43.8	*18-Nov-21 14:47:27	23.1	44.9	18-Nov-21 16:27:27	21.2	48.6	18-Nov-21 18:07:27	21.2	48.9
*18-Nov-21 09:48:27	20.9	50.5	*18-Nov-21 11:28:27	22.2	47.6	*18-Nov-21 13:08:27	23.1	44.0	*18-Nov-21 14:48:27	23.1	44.8	18-Nov-21 16:28:27	21.2	48.8	18-Nov-21 18:08:27	21.2	48.8
*18-Nov-21 09:49:27	21.0	50.6	*18-Nov-21 11:29:27	22.2	47.0	*18-Nov-21 13:09:27	23.1	44.5	*18-Nov-21 14:49:27	23.1	44.8	18-Nov-21 16:29:27	21.1	48.7	18-Nov-21 18:09:27	21.2	48.8
*18-Nov-21 09:50:27	21.0	50.5	*18-Nov-21 11:30:27	22.2	47.2	*18-Nov-21 13:10:27	23.2	43.2	*18-Nov-21 14:50:27	23.1	45.0	18-Nov-21 16:30:27	21.1	48.8	18-Nov-21 18:10:27	21.2	48.8
*18-Nov-21 09:51:27	21.0	50.5	*18-Nov-21 11:31:27	22.2	47.0	*18-Nov-21 13:11:27	23.2	45.1	*18-Nov-21 14:51:27	23.1	44.7	18-Nov-21 16:31:27	21.0	48.8	18-Nov-21 18:11:27	21.2	48.7
*18-Nov-21 09:52:27	21.0	50.5	*18-Nov-21 11:32:27	22.2	47.1	*18-Nov-21 13:12:27	23.2	44.3	*18-Nov-21 14:52:27	23.2	44.7	18-Nov-21 16:32:27	21.0	49.1	18-Nov-21 18:12:27	21.2	48.8
*18-Nov-21 09:53:27	21.0	50.5	*18-Nov-21 11:33:27	22.1	46.3	*18-Nov-21 13:13:27	23.1	42.4	*18-Nov-21 14:53:27	23.2	44.3	18-Nov-21 16:33:27	20.9	49.3	18-Nov-21 18:13:27	21.2	48.8
*18-Nov-21 09:54:27	21.0	50.6	*18-Nov-21 11:34:27	22.1	46.5	*18-Nov-21 13:14:27	23.1	43.4	*18-Nov-21 14:54:27	23.0	45.0	18-Nov-21 16:34:27	20.8	49.2	18-Nov-21 18:14:27	21.2	48.7
*18-Nov-21 09:55:27	21.0	50.6	*18-Nov-21 11:35:27	22.1	46.1	*18-Nov-21 13:15:27	23.1	43.0	*18-Nov-21 14:55:27	23.1	44.5	18-Nov-21 16:35:27	20.8	49.2	18-Nov-21 18:15:27	21.2	48.8
*18-Nov-21 09:56:27	21.0	50.5	*18-Nov-21 11:36:27	22.1	45.7	*18-Nov-21 13:16:27	23.0	43.9	*18-Nov-21 14:56:27	23.1	45.4	18-Nov-21 16:36:27	20.8	49.3	18-Nov-21 18:16:27	21.2	48.7
*18-Nov-21 09:57:27	21.0	50.5	*18-Nov-21 11:37:27	22.0	45.9	*18-Nov-21 13:17:27	23.1	43.6	*18-Nov-21 14:57:27	23.0	45.4	18-Nov-21 16:37:27	20.8	49.5	18-Nov-21 18:17:27	21.2	48.7
*18-Nov-21 09:58:27	21.0	50.5	*18-Nov-21 11:38:27	22.0	47.0	*18-Nov-21 13:18:27	23.1	44.5	*18-Nov-21 14:58:27	23.0	45.7	18-Nov-21 16:38:27	20.8	49.5	18-Nov-21 18:18:27	21.1	48.8
*18-Nov-21 09:59:27	21.0	50.6	*18-Nov-21 11:39:27	22.0	46.8	*18-Nov-21 13:19:27	23.1	43.9	*18-Nov-21 14:59:27	23.0	45.3	18-Nov-21 16:39:27	20.8	49.5	18-Nov-21 18:19:27	21.2	48.7
*18-Nov-21 10:00:27	21.1	50.6	*18-Nov-21 11:40:27	22.0	46.4	*18-Nov-21 13:20:27	23.2	44.3	*18-Nov-21 15:00:27	23.0	45.5	18-Nov-21 16:40:27	20.9	49.5	18-Nov-21 18:20:27	21.2	48.7
*18-Nov-21 10:01:27	21.2	56.1	*18-Nov-21 11:41:27	21.9	44.8	*18-Nov-21 13:21:27	23.2	42.6	*18-Nov-21 15:01:27	22.9	45.9	18-Nov-21 16:41:27	20.9	49.6	18-Nov-21 18:21:27	21.2	48.8
*18-Nov-21 10:02:27	21.2	55.3	*18-Nov-21 11:42:27	22.0	45.8	*18-Nov-21 13:22:27	23.2	43.7	*18-Nov-21 15:02:27	23.0	46.1	18-Nov-21 16:42:27	21.0	49.7	18-Nov-21 18:22:27	21.1	48.7
*18-Nov-21 10:03:27	21.3	51.0	*18-Nov-21 11:43:27	22.0	46.7	*18-Nov-21 13:23:27	23.2	42.7	*18-Nov-21 15:03:27	23.0	46.0	18-Nov-21 16:43:27	21.0	49.7	18-Nov-21 18:23:27	21.1	48.7
*18-Nov-21 10:04:27	21.4	53.3	*18-Nov-21 11:44:27	22.0	45.7	*18-Nov-21 13:24:27	23.2	42.5	*18-Nov-21 15:04:27	23.0	45.7	18-Nov-21 16:44:27	21.0	49.7	18-Nov-21 18:24:27	21.1	48.7
*18-Nov-21 10:05:27	21.4	51.9	*18-Nov-21 11:45:27	21.9	45.4	*18-Nov-21 13:25:27	23.2	42.1	*18-Nov-21 15:05:27	23.0	45.7	18-Nov-21 16:45:27	21.0	49.7	18-Nov-21 18:25:27	21.2	48.7
*18-Nov-21 10:06:27	21.4	50.9	*18-Nov-21 11:46:27	21.9	46.1	*18-Nov-21 13:26:27	23.2	42.4	*18-Nov-21 15:06:27	23.0	45.9	18-Nov-21 16:46:27	21.1	49.7	18-Nov-21 18:26:27	21.1	48.7
*18-Nov-21 10:07:27	21.3	50.8	*18-Nov-21 11:47:27	22.0	46.5	*18-Nov-21 13:27:27	23.2	42.1	*18-Nov-21 15:07:27	23.0	45.9	18-Nov-21 16:47:27	21.1	49.7	18-Nov-21 18:27:27	21.1	48.7
*18-Nov-21 10:08:27	21.4	50.6	*18-Nov-21 11:48:27	22.0	46.7	*18-Nov-21 13:28:27	23.2	41.6	*18-Nov-21 15:08:27	23.0	45.7	18-Nov-21 16:48:27	21.1	49.8	18-Nov-21 18:28:27	21.1	48.7
*18-Nov-21 10:09:27	21.4	50.6	*18-Nov-21 11:49:27	22.0	47.1	*18-Nov-21 13:29:27	23.3	41.8	*18-Nov-21 15:09:27	22.9	45.7	18-Nov-21 16:49:27	21.1	49.7	18-Nov-21 18:29:27	21.1	48.7
*18-Nov-21 10:10:27	21.4	50.7	*18-Nov-21 11:50:27	21.9	47.1	*18-Nov-21 13:30:27	23.3	41.6	*18-Nov-21 15:10:27	23.0	45.9	18-Nov-21 16:50:27	21.2	49.8	18-Nov-21 18:30:27	21.1	48.7
*18-Nov-21 10:11:27	21.4	50.8	*18-Nov-21 11:51:27	22.0	47.0	*18-Nov-21 13:31:27	23.3	42.4	*18-Nov-21 15:11:27	23.0	45.8	18-Nov-21 16:51:27	21.2	49.7	18-Nov-21 18:31:27	21.1	48.6
*18-Nov-21 10:12:27	21.4	52.1	*18-Nov-21 11:52:27	22.0	46.5	*18-Nov-21 13:32:27	23.3	42.2	*18-Nov-21 15:12:27	23.0	45.3	18-Nov-21 16:52:27	21.1	49.7	18-Nov-21 18:32:27	21.1	48.6
*18-Nov-21 10:13:27	21.4	50.7	*18-Nov-21 11:53:27	22.0	48.3	*18-Nov-21 13:33:27	23.2	42.3	*18-Nov-21 15:13:27	22.9	45.7	18-Nov-21 16:53:27	21.2	49.8	18-Nov-21 18:33:27	21.1	48.7
*18-Nov-21 10:14:27	21.4	50.8	*18-Nov-21 11:54:27	22.0	47.3	*18-Nov-21 13:34:27	23.2	42.7	*18-Nov-21 15:14:27	22.9	45.8	18-Nov-21 16:54:27	21.2	49.8	18-Nov-21 18:34:27	21.0	48.7
*18-Nov-21 10:15:27	21.5	50.8	*18-Nov-21 11:55:27	22.1	47.1	*18-Nov-21 13:35:27	23.2	42.2	*18-Nov-21 15:15:27	22.9	45.8	18-Nov-21 16:55:27	21.2	49.8	18-Nov-21 18:35:27	21.1	48.6
*18-Nov-21 10:16:27	21.4	50.9	*18-Nov-21 11:56:27	22.1	47.3	*18-Nov-21 13:36:27	23.3										



































































DATE TIME	°C	%RH	DATE TIME	°C	%RH	DATE TIME	°C	%RH	DATE TIME	°C	%RH	DATE TIME	°C	%RH
30-Nov-21 11:25:27	20.8	36.9	30-Nov-21 13:05:27	20.5	35.3									
30-Nov-21 11:26:27	20.9	36.3	30-Nov-21 13:06:27	20.4	35.2									
30-Nov-21 11:27:27	20.9	36.2	30-Nov-21 13:07:27	20.4	35.3									
30-Nov-21 11:28:27	20.9	36.3	30-Nov-21 13:08:27	20.5	35.3									
30-Nov-21 11:29:27	20.9	36.2	30-Nov-21 13:09:27	20.4	35.3									
30-Nov-21 11:30:27	20.9	36.1	30-Nov-21 13:10:27	20.5	35.3									
30-Nov-21 11:31:27	21.0	36.1	30-Nov-21 13:11:27	20.5	35.2									
30-Nov-21 11:32:27	20.9	36.1	30-Nov-21 13:12:27	20.4	35.2									
30-Nov-21 11:33:27	20.9	36.1	30-Nov-21 13:13:27	20.4	35.3									
30-Nov-21 11:34:27	21.0	36.0	30-Nov-21 13:14:27	20.4	35.3									
30-Nov-21 11:35:27	21.0	35.9	30-Nov-21 13:15:27	20.4	35.4									
30-Nov-21 11:36:27	21.0	36.0	30-Nov-21 13:16:27	20.4	35.3									
30-Nov-21 11:37:27	21.0	36.3	30-Nov-21 13:17:27	20.4	35.3									
30-Nov-21 11:38:27	21.0	35.7	30-Nov-21 13:18:27	20.4	35.2									
30-Nov-21 11:39:27	21.0	33.3	30-Nov-21 13:19:27	20.4	35.3									
30-Nov-21 11:40:27	20.8	33.6	30-Nov-21 13:20:27	20.4	35.3									
30-Nov-21 11:41:27	20.7	34.7	30-Nov-21 13:21:27	20.4	35.2									
30-Nov-21 11:42:27	20.7	34.9	30-Nov-21 13:22:27	20.4	35.2									
30-Nov-21 11:43:27	20.7	35.0	30-Nov-21 13:23:27	20.4	35.2									
30-Nov-21 11:44:27	20.7	35.0	30-Nov-21 13:24:27	20.4	35.3									
30-Nov-21 11:45:27	20.7	35.0	30-Nov-21 13:25:27	20.4	35.2									
30-Nov-21 11:46:27	20.7	35.0	30-Nov-21 13:26:27	20.4	35.3									
30-Nov-21 11:47:27	20.7	35.0	30-Nov-21 13:27:27	20.4	35.2									
30-Nov-21 11:48:27	20.7	34.9	30-Nov-21 13:28:27	20.4	35.3									
30-Nov-21 11:49:27	20.7	34.9	30-Nov-21 13:29:27	20.4	35.3									
30-Nov-21 11:50:27	20.7	34.9	30-Nov-21 13:30:27	20.4	35.3									
30-Nov-21 11:51:27	20.8	34.9	30-Nov-21 13:31:27	20.4	35.3									
30-Nov-21 11:52:27	20.8	34.8	30-Nov-21 13:32:27	20.4	35.3									
30-Nov-21 11:53:27	20.7	34.7	30-Nov-21 13:33:27	20.4	35.3									
30-Nov-21 11:54:27	20.8	34.8	30-Nov-21 13:34:27	20.4	35.3									
30-Nov-21 11:55:27	20.7	34.9	30-Nov-21 13:35:27	20.4	35.5									
30-Nov-21 11:56:27	20.7	34.8	30-Nov-21 13:36:27	20.4	35.4									
30-Nov-21 11:57:27	20.7	34.7	30-Nov-21 13:37:27	20.4	35.4									
30-Nov-21 11:58:27	20.7	34.7	30-Nov-21 13:38:27	20.4	35.4									
30-Nov-21 11:59:27	20.7	34.8	30-Nov-21 13:39:27	20.3	35.4									
30-Nov-21 12:00:27	20.6	34.8	30-Nov-21 13:40:27	20.4	35.4									
30-Nov-21 12:01:27	20.6	34.8	30-Nov-21 13:41:27	20.3	35.4									
30-Nov-21 12:02:27	20.6	34.8	30-Nov-21 13:42:27	20.4	35.4									
30-Nov-21 12:03:27	20.6	34.8	30-Nov-21 13:43:27	20.3	35.4									
30-Nov-21 12:04:27	20.6	34.8	30-Nov-21 13:44:27	20.3	35.4									
30-Nov-21 12:05:27	20.6	34.9	30-Nov-21 13:45:27	20.4	35.4									
30-Nov-21 12:06:27	20.6	34.9	30-Nov-21 13:46:27	20.3	35.5									
30-Nov-21 12:07:27	20.6	34.9	30-Nov-21 13:47:27	20.3	35.4									
30-Nov-21 12:08:27	20.6	35.0	30-Nov-21 13:48:27	20.4	35.6									
30-Nov-21 12:09:27	20.6	34.9	30-Nov-21 13:49:27	20.4	35.7									
30-Nov-21 12:10:27	20.6	34.9	30-Nov-21 13:50:27	20.4	35.7									
30-Nov-21 12:11:27	20.5	34.9	30-Nov-21 13:51:27	20.4	35.7									
30-Nov-21 12:12:27	20.6	35.0	30-Nov-21 13:52:27	20.4	35.8									
30-Nov-21 12:13:27	20.5	35.0	30-Nov-21 13:53:27	20.4	35.7									
30-Nov-21 12:14:27	20.5	35.0	30-Nov-21 13:54:27	20.4	35.7									
30-Nov-21 12:15:27	20.6	35.0	30-Nov-21 13:55:27	20.4	35.7									
30-Nov-21 12:16:27	20.5	35.1	30-Nov-21 13:56:27	20.4	35.6									
30-Nov-21 12:17:27	20.5	35.1	30-Nov-21 13:57:27	20.4	35.4									
30-Nov-21 12:18:27	20.5	35.2	30-Nov-21 13:58:27	20.5	35.4									
30-Nov-21 12:19:27	20.5	35.2	30-Nov-21 13:59:27	20.5	35.6									
30-Nov-21 12:20:27	20.5	35.2	30-Nov-21 14:00:27	20.5	35.5									
30-Nov-21 12:21:27	20.5	35.1	30-Nov-21 14:01:27	20.5	35.6									
30-Nov-21 12:22:27	20.5	35.2	30-Nov-21 14:02:27	20.5	35.4									
30-Nov-21 12:23:27	20.5	35.1	30-Nov-21 14:03:27	20.5	35.4									
30-Nov-21 12:24:27	20.6	35.3	30-Nov-21 14:04:27	20.6	35.5									
30-Nov-21 12:25:27	20.5	35.2	30-Nov-21 14:05:27	20.5	35.3									
30-Nov-21 12:26:27	20.6	35.1	30-Nov-21 14:06:27	20.5	35.3									
30-Nov-21 12:27:27	20.6	35.2	30-Nov-21 14:07:27	20.5	35.3									
30-Nov-21 12:28:27	20.5	35.1	30-Nov-21 14:08:27	20.5	35.3									
30-Nov-21 12:29:27	20.5	35.1	30-Nov-21 14:09:27	20.6	35.3									
30-Nov-21 12:30:27	20.5	35.0	30-Nov-21 14:10:27	20.6	35.3									
30-Nov-21 12:31:27	20.5	35.1	30-Nov-21 14:11:27	20.6	35.2									
30-Nov-21 12:32:27	20.6	35.1	30-Nov-21 14:12:27	20.6	35.1									
30-Nov-21 12:33:27	20.5	35.1	30-Nov-21 14:13:27	20.6	35.2									
30-Nov-21 12:34:27	20.5	35.1	30-Nov-21 14:14:27	20.6	35.2									
30-Nov-21 12:35:27	20.5	35.1	30-Nov-21 14:15:27	20.6	35.2									
30-Nov-21 12:36:27	20.5	35.2	30-Nov-21 14:16:27	20.6	35.1									
30-Nov-21 12:37:27	20.5	35.0	30-Nov-21 14:17:27	20.6	35.1									
30-Nov-21 12:38:27	20.5	35.1	30-Nov-21 14:18:27	20.7	35.1									
30-Nov-21 12:39:27	20.5	35.2	30-Nov-21 14:19:27	20.7	35.0									
30-Nov-21 12:40:27	20.5	35.1	30-Nov-21 14:20:27	20.7	35.0									
30-Nov-21 12:41:27	20.5	35.1	30-Nov-21 14:21:27	20.7	35.0									
30-Nov-21 12:42:27	20.5	35.1	30-Nov-21 14:22:27	20.7	34.9									
30-Nov-21 12:43:27	20.5	35.1	30-Nov-21 14:23:27	20.7	34.9									
30-Nov-21 12:44:27	20.5	35.2	30-Nov-21 14:24:27	20.7	34.9									
30-Nov-21 12:45:27	20.5	35.2	30-Nov-21 14:25:27	20.7	34.9									
30-Nov-21 12:46:27	20.5	35.2	30-Nov-21 14:26:27	20.7	34.8									
30-Nov-21 12:47:27	20.5	35.3	30-Nov-21 14:27:27	20.8	34.8									
30-Nov-21 12:48:27	20.5	35.3	30-Nov-21 14:28:27	20.8	34.7									
30-Nov-21 12:49:27	20.5	35.2	30-Nov-21 14:29:27	20.7	34.8									
30-Nov-21 12:50:27	20.5	35.3	30-Nov-21 14:30:27	20.8	34.9									
30-Nov-21 12:51:27	20.5	35.2	30-Nov-21 14:31:27	20.8	34.7									
30-Nov-21 12:52:27	20.5	35.3	30-Nov-21 14:32:27	20.8	34.8									
30-Nov-21 12:53:27	20.5	35.3	30-Nov-21 14:33:27	20.8	34.8									
30-Nov-21 12:54:27	20.5	35.2	30-Nov-21 14:34:27	20.8	34.7									
30-Nov-21 12:55:27	20.5	35.3	30-Nov-21 14:35:27	20.8	34.6									
30-Nov-21 12:56:27	20.5	35.2	30-Nov-21 14:36:27	20.8	34.7									
30-Nov-21 12:57:27	20.4	35.3	30-Nov-21 14:37:27	20.9	34.6									
30-Nov-21 12:58:27	20.5	35.3												
30-Nov-21 12:59:27	20.5	35.3												
30-Nov-21 13:00:27	20.4	35.2												
30-Nov-21 13:01:27	20.5	35.3												
30-Nov-21 13:02:27	20.5	35.2												
30-Nov-21 13:03:27	20.4	35.3												
30-Nov-21 13:04:27	20.5	35.4												

## Manufacturer's Certificate of Validation

<b>Certificate No.</b>	<b>Date</b>	<b>Page</b>
VC21050014-17	June 06,2021	1 of 1

tempmate GmbH certifies that the products mentioned on the following page(s) have been thoroughly tested, validated and met performance accuracy specifications over the stated ranges.

### Reference Instrumentation

Model	Calibration Date	Accuracy	Serial Number	CNAS Certificate Number
XORTS-120A	August 20,2020	± 0.01°C	160408-1	H2020-3052842
ATL60	August 20,2020	± 0.015°C	61080016	H2020-3052841
C4-180Pro	August 20,2020	± 1%RH	54260019170010	H2020-3052843

All reference instruments are traceable calibrated by CNAS accredited laboratories.

### Environment Conditions

<b>Air Temperature</b>	<b>25.5 °C</b>	<b>Relative Humidity</b>	<b>53.0%RH</b>
------------------------	----------------	--------------------------	----------------

### Validation Information

Validation Points	Tolerance	Result	Validation Points	Tolerance	Result
+60 °C	± 0.3 °C	PASS	70%RH	± 3%RH	PASS
+40 °C	± 0.3 °C	PASS	60%RH	± 3%RH	PASS
+25 °C	± 0.3 °C	PASS	50%RH	± 3%RH	PASS
0 °C	± 0.3 °C	PASS	40%RH	± 3%RH	PASS
-5 °C	± 0.3 °C	PASS	30%RH	± 3%RH	PASS
-25 °C	± 0.3 °C	PASS	20%RH	± 3%RH	PASS

Temp Accuracy ± 0.3 °C

Hum Accuracy 20~80%RH(25 °C: ± 3%RH) Others± 5%RH

### Product Information

<b>Product Name</b>	<b>Serial Number</b>	<b>Production Date</b>	<b>Expiry Date</b>	<b>Quantity</b>
tempmate <sup>®</sup> -S2TH	S2H21050014-17	06/2021	06/2023	1 pcs

This certificate was automatically generated and is valid without signature for two years.

The calibration of the digital sensor is valid for 2 years from the date of production, considering the device is handled/stored under recommended conditions. V05/2021